16048D Test Leads



Terminal Connector: 4-Terminal Pair, BNC Cable Length (approx.):

1.89 m (from connector to cable tip)

Cable Tip: BNC (male)
Weight (approx.): 460 g

Additional Error: For detailed information, refer to the measurement instrument's specifications.

Description: The test leads extend the measurement port with a 4-Terminal Pair configuration. It is provided with a BNC female connector board to allow the attachment of user-fabricated test fixtures.

 $\textbf{Applicable Instruments:}\ 4263B,\ 4268A,\ 4279A^*,$

4284A*1, 4285A, 4288A, E4980A, E4981A

* denotes the instrument is obsolete.

1. Requires Option 006.

Frequency: DC to 30 MHz

Maximum Voltage: ±40 V peak max (AC+DC)

Operating Temperature: 0°C to 55°C

Furnished Accessories:

| Description | P/N | Qty. |
|------------------------------|-------------|------|
| Terminal Board with BNC(f)x4 | 16032-60001 | 1 |
| Operation Manual | 16048-90031 | 1 |

Compensation and Measurement: Cable length compensation is recommended before measurement. Set the instrument's cable length compensation function to 2 m.

16048E Test Leads



Terminal Connector: 4-Terminal Pair, BNC

Cable Length (approx.): 3.8 m(from connector to cable tip)
Cable Tip: BNC (male)

Weight (approx.): 690 g

Additional Error: For detailed information, refer to the measurement instrument's specifications.

Description: The test leads extend the measurement port with a 4-Terminal Pair configuration. It is provided with a BNC female connector board to allow the attachment of user-fabricated test fixtures.

Applicable Instruments: 4263B, 4284A*1, E4980A

* denotes the instrument is obsolete.

1. Requires Option 006.

Frequency: DC to 2 MHz

Maximum Voltage: ±40 V peak max (AC+DC)

Operating Temperature: 0°C to 55°C

Furnished Accessories:

| Description | P/N | Qty. |
|------------------------------|-------------|------|
| Terminal Board with BNC(f)x4 | 16032-60001 | 1 |
| Operation Manual | 16048-90041 | 1 |

Compensation and Measurement: Cable length compensation is recommended before measurement. Set the instrument's cable length compensation function to 4 m.