

PV-ISOTEST

Rel. 2.00 - 16/11/20

Safety tester for Insulation up to 1500VDC in PV plants

Page 1 of 2

2. ELECTRICAL SPECIFICATIONS

Accuracy is calculated as \pm [% readings + (no. of digits) * resolution] at 23 ° C \pm 5 ° C, relative humidity <80%RH

DC Voltage		
Range (V)	Resolution (V)	Accuracy
3 ÷ 1500	1	± (1.0%rdg + 2dgt)

AC TRMS Voltage		
Range (V)	Resolution (V)	Accuracy
3 ÷ 1000	1	± (1.0%rdg + 3dgt)

Frequency range: 42.5 ÷ 69Hz; Voltage zeroed for measured values <3V

Insulation Resistance (M Ω) – DUAL Mode				
Test voltage DC [V]	Range [MΩ]	Resolution [MΩ]	Accuracy (*)	
250, 500, 1000, 1500	0.1 ÷ 0.99	0.01		
	1.0 ÷ 19.9	0.1	±(5%rdg + 5dgt)	
	20 ÷ 100	1	, , ,	

^(*) Accuracy indicatec for VPN ≥240V, Rfault≥10Ω. Accuracy of Rp and R(+) not declared if R(+)≥ 0.2MΩ and R(-) <0.2MΩ

Accuracy of Rp and R(-) not declared if R(+) < $0.2M\Omega$ and R(-) $\ge 0.2M\Omega$

Open voltage <1.25 x nominal test voltage Short circuit current <15mA (peak) for each test voltage

Nominal measured current >1mA on R = $1k\Omega \times Vnom \text{ (with VPN, VPE, VNE= 0)}$

Insulation Resistance (MΩ) –TIMER Mode				
Test voltage DC [V]	Range [MΩ]	Resolution [MΩ]	Accuracy	
250 500 1000 1500	0.01 ÷ 9.99	0.01	1/F 00/rdg : Edgt)	
250, 500, 1000, 1500	10.0 ÷ 99.9	0.1	±(5.0%rdg+ 5dgt)	

Open voltage <1.25 x nominal test voltage
Short circuit current <15mA (peak) for each test voltage

Nominal measured current >1mA on R = 1k Ω x Vnom (with VPN, VPE, VNE= 0)

Setting timer: 3s ÷ 999s

Continuity of protection conductors (RPE)			
Range [Ω]	Resolution [Ω]	Accuracy	
$0.00 \div 9.99$	0.01		
10.0 ÷ 99.9	0.1	±(2%rdg + 2dgt)	
100 ÷ 1999	1]	

Test current: >200mA DC up to 5Ω (included cables), Resolution 1mA, Accuracy \pm (5.0%rdg + 5dgt)

Open voltage $4 < V_0 < 10V$

GFL (Ground Fault Locator) function				
Test voltage DC [V]	Range [MΩ]	Resolution [M Ω]	Accuracy (*)	Position accuracy
	0.1 ÷ 0.99	0.01		
250, 500, 1000, 1500	1.0 ÷ 19.9	0.1	\pm (5%rdg + 5dgt)	± 1module
	20 ÷ 100	1		

^(*)Accuracy indicatec for VPN ≥240V, Rfault≥10Ω. Accuracy of Rp and R(+) not declared if R(+)≥ 0.2MΩ and R(-) <0.2MΩ

Accuracy of Rp and R(-) not declared if R(+) < $0.2M\Omega$ and R(-) $\geq 0.2M\Omega$

Open voltage <1.25 x nominal test voltage
Short circuit current <15mA (peak) for each test voltage

Nominal measured current >1mA on R = $1k\Omega \times V$ nom (with VPN, VPE, VNE= 0) The GFL function allows obtaining correct results with the following conditions:

> Test carried out with Vtest ≥Vnom on a single string disconnected from the inverter, from possible arresters and from earth connections

> Test performed upstream of any blocking diodes

> Single fault of low insulation located at any position in the string

> Insulation resistance of the single fault <0.1M Ω

Environmental conditions similar to those in which the fault was reported

HT ITALIA SRL

WWW. Viadalia Bertin 6 1480 5 aere of An Italy

Tel: +39-0546-621002 - Fax: +39-0546-621144 email: export@htitalia.it - web: http://www.ht-instruments.com



PV-ISOTEST

Rel. 2.00 - 16/11/20

Safety tester for Insulation up to 1500VDC in PV plants

Page 2 of 2

2. GENERAL SPECIFICATIONS

DISPLAY AND MEMORY:

Features: graphic COG 128x128pxl with backlight

Memory: max 999 test

POWER SUPPLY:

Battery type: 6x1.5V alkaline batteries type AA LR06 or

6x1.1V rechargeable batteries type AA LR06

Battery life: approx. 500 tests (for each functions)

Auto Power OFF: after 5 minutes o idleness

OUTPUT INTERFACE

PC communication port: optical/USB

MECHANICAL SPECIFICATIONS

Dimensions (L x W x H): 235 x 165 x 75mm

Weight (batteries included): 1.2kg Mechanical protection: 1P40

ENVIRONMENTAL CONDITIONS:

Reference temperature: $23^{\circ}\text{C} \pm 5^{\circ}\text{C}$ Working temperature: $0^{\circ}\text{C} \div 40^{\circ}\text{C}$ Working humidity: $<80^{\circ}\text{RH}$ Storage temperature: $-10^{\circ}\text{C} \div 60^{\circ}\text{C}$ Storage humidity: $<80^{\circ}\text{RH}$ Max height of use: 2000°m

REFERENCE GUIDELINES:

Instrument's safety: IEC/EN61010-1, IEC/EN61010-2-030

IEC/EN61010-2-033, IEC/EN61010-2-034

EMC: IEC/EN61326-1
Safety of measurement accessories: IEC/EN61010-031
General: IEC/EN62446

 $\begin{array}{lll} \text{Measurement M}\Omega & \text{IEC/EN 61557-2} \\ \text{Measurement RPE:} & \text{IEC/EN 61557-4} \\ \text{Insulation:} & \text{double insulation} \end{array}$

Pollution degree: 2

Overvoltage category: CAT III 1500V DC, CAT III 1000V AC

Max 1500V DC, 1000VAC between inputs

This instrument complies with the requirements of the European Low Voltage Directives 2014/35/EU (LVD) and EMC 2014/30/EU

This instrument satisfies the requirements of 2011/65/EU (RoHS) directive and 2012/19/EU (WEEE) directive

Tel: +39-0546-621002 - Fax: +39-0546-621144 email: export@htitalia.it - web: http://www.ht-instruments.com